ANIOSYME PRIME

Nettoyant et désinfectant de l'instrumentation



LABORATOIRES ANIOS

An Ecolab Company

750ml

UN23670750MCAR-201202

ANIOSYME PRIME

(EN) Cleaning and disinfection of the instrumentation

INDICATIONS: Pre-treatment of surgical instrumentation in foam: during a dry mode of transport or during storage of soiled instruments before machine or manual washing. Maintains an adequate level of humidity to avoid blood and other organic materials drying. **COMPOSITION:** Quaternary ammonium carbonate, non-ionic surfactants, enzymatic complex (protease, lipase, and amylase), sequestering agent and excipients. **MICROBIOLOGICAL PROPERTIES:** Bactericidal* according to EN 13727/EN 14561 in 15 min. Yeasticidal* according to EN 13624/EN 14562 in 15 min. Tuberculocidal* according to EN 14348/EN 14563 in 30 min. Mycobactericidal* according to EN 14348/ EN 14563 in 60 min. Active against enveloped viruses* according to EN 14476/EN 17111 in 15 min. *=dirty conditions. **INSTRUCTIONS FOR USE:** Ready-to-use solution. Beforehand, spray the bottom of the soaking tray with Aniosyme Prime and place instruments collected. Re-cover all the instruments with foam. Cover the tray with the lid. Minimum contact time in the system: 15 minutes. For tuberculocidal activity, allow 30 minutes contact time. Thoroughly rinse treated medical devices with tap water to remove all traces of product and organic soiling before machine or manual cleaning. Not suitable for instruments made of zinc or iron. Scientific dossier (efficacy, compatibility, etc.) available on request. **PRECAUTIONS FOR USE:** WARNING. Causes serious eye irritation. Wear eye/face protective equipment. Remove the product and its container by following all the warnings of use. Keep out of the reach of children. Anti-Poison Center Tel: 210 7793777 (Greece), 1401 (Cyprus). Reserved for professional use. Medical Device. Made in France.

(GR) Καθαρισμός και απολύμανση εργαλείων

ΕΝΔΕΙΞΕΙΣ: Προεπεξεργασία ιατρικών και χειρουργικών εργαλείων με αφρό: κατά την μεταφορά χωρίς υγρό μέσο ή κατά την αποθήκευση ακάθαρτων εργαλείων πριν από το πλύσιμο σε μηχάνημα ή στο χέρι. Διατηρεί επαρκή υγρασία για την αποφυγή της ξήρανσης του αίματος και άλλων οργανικών ουσιών. **ΣΥΝΘΕΣΗ:** Ανθρακικό τεταρτοταγές αμμώνιο, μη ιονικά επιφανειοδραστικά, σύμπλοκο ενζύμων (πρωτεάση, λιπάση και αμυλάση), συμπλοκοποιητής και έκδοχα. **ΜΙΚΡΟΒΙΟΛΟΓΙΚΕΣ ΙΔΙΟΤΗΤΕΣ:** Βακτηριοκτόνο* σύμφωνα με την EN 13727/EN 14561 σε 15 λεπτά. Ζυμομυκητοκτόνο* σύμφωνα με την EN 13624/ EN 14562 σε 15 λεπτά. Φυματιοκτόνο* σύμφωνα με την EN 14348/EN 14563 σε 30 λεπτά. Μυκοβακτηριοκτόνο* σύμφωνα με την EN 14348/ EN 14563 σε 60 λεπτά. Δραστικό έναντι των ιών με μανδύα* σύμφωνα με την EN 14476/EN 17111 σε 15 λεπτά. *= ρυπαρές συνθήκες. **ΤΡΟΠΟΣ ΧΡΗΣΗΣ:** Διάλυμα έτοιμο για χρήση. Αρχικά, ψεκάστε τον πυθμένα του δοχείου διαβροχής με Aniosyme Prime και τοποθετήστε στο δοχείο τα εργαλεία που μαζέψατε. Καλύψτε όλα τα εργαλεία με αφρό. Τοποθετήστε το κάλυμμα στο δοχείο. Ελάχιστος χρόνος επαφής: 15 λεπτά. Για δράση φυματιοκτόνου, τηρείτε χρόνο επαφής 30 λεπτών. Ξεπλύνετε τις ιατρικές συσκευές που υποβλήθηκαν σε καθαρισμό με άφθονο νερό της βρύσης για να απομακρύνετε κάθε ίχνος προϊόντος και οργανικών ακαθαρσιών πριν προβείτε σε καθαρισμό χειροκίνητα ή με μηχάνημα. Δεν είναι κατάλληλο για εργαλεία από ψευδάργυρο ή σίδηρο. Επιστημονικός φάκελος (αποτελεσματικότητα, συμβατότητα...) διαθέσιμος μετά από αίτημα. **ΠΡΟΦΥΛΑΞΕΙΣ ΚΑΤΑ ΤΗ ΧΡΗΣΗ:** ΠΡΟΣΟΧΗ. Προκαλεί σοβαρό οφθαλμικό ερεθισμό. Να φοράτε μέσα ατομικής προστασίας για τα μάτια/πρόσωπο. Διάθεση του προϊόντος και του δοχείου του λαμβάνοντας όλες τις προφυλάξεις χρήσης. Μακρυά από παιδιά. Τ ηλ. Κέντ. Δηλητηριάσεων: 210 7793777 (Ελλάδα), 1401 (Κύπρος). Αποκλειστικά για επαγγελματική χρήση. Ιατροτεχνολογικό προϊόν. Παράγεται στην Γαλλία.

CE
0459





UFI: EW4X-YQN6
2F0F-HMYA

OB2367890750MCAV-210913

LABORATOIRES ANIOS
An Ecolab Company
Laboratoires ANIOS
1, rue de l'Espoir
59260 Lezennes - FRANCE
Tél : +33 3 20 67 67 67
Fax : +33 3 20 67 67 68
www.anios.com

750ml

3 597610 409707

LOT